

## PATENT

## IN THE UNITED STATES PATENT AND TRADEMARK OFFICE

In re Application of:	Lixiao Wang
Application No.:	08/685338
Filed:	July 23, 1996
For:	High Compliance, High Strength Catheter Balloons Useful for Treatment of Gastrointestinal Lesions
Examiner:	Cris L. Rodrigues
Group Art Unit:	3763

Box Non-Fee Amendment  
Assistant Commissioner for Patents  
Washington, D.C. 20231

Docket No.: S63.2B-5902-US

**Amendment**

This amendment is in response to the office action mailed January 28, 2003.

Please amend the application as follows:

**In The Claims:**

43. (Amended) A balloon for a medical device characterized by a burst pressure of at least 9 atmospheres, a diameter at 3 atmospheres of about 2 mm or more, and an average compliance over the range of from 3 atmospheres to burst of at least 3% per atmosphere, wherein the balloon is formed from at least two concentric layers of different thermoplastic polymers.

*Attached hereto is a marked-up version of the changes made to the specification and claims by the current amendment. The attached page is captioned "Version with markings to show changes made."*

**Remarks**

Claim 43 has been amended according to the requirements in the official action